

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
 Data File : VY010080.D
 Acq On : 17 Aug 2022 19:46
 Operator : KP/MD
 Sample : N4124-04
 Misc : 12.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SSP4

Quant Time: Aug 18 08:23:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 17 05:12:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	68230	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	97951	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	75706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	22576	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	19458	28.522	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	57.040%
35) Dibromofluoromethane	7.710	113	17815	28.046	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	56.100%
50) Toluene-d8	10.179	98	110690	47.100	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.200%
62) 4-Bromofluorobenzene	12.477	95	23104	27.990	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	55.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081722\
Data File : VY010080.D
Acq On : 17 Aug 2022 19:46
Operator : KP/MD
Sample : N4124-04
Misc : 12.06g/5.0mL/MSVOA_Y/SoIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SSP4

Quant Time: Aug 18 08:23:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 17 05:12:02 2022
Response via : Initial Calibration

