

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081723\
 Data File : VY015172.D
 Acq On : 17 Aug 2023 15:17
 Operator : SY/MD
 Sample : 04046-04
 Misc : 3.73g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-07(0-5)

Quant Time: Aug 18 00:54:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	126926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	221591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	223962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	113578	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	85035	55.811	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.620%
35) Dibromofluoromethane	7.710	113	71013	49.814	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.620%
50) Toluene-d8	10.179	98	269494	50.783	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.560%
62) 4-Bromofluorobenzene	12.483	95	101347	52.287	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081723\
Data File : VY015172.D
Acq On : 17 Aug 2023 15:17
Operator : SY/MD
Sample : 04046-04
Misc : 3.73g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-07(0-5)

Quant Time: Aug 18 00:54:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

