



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081723\
Data File : VY015160.D
Acq On : 17 Aug 2023 09:25
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Title : SW846 8260
Last Update : Mon Aug 14 12:43:26 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	22373	PASS
75	95	30	60	55.0	60173	PASS
95	95	100	100	100.0	109400	PASS
96	95	5	9	6.7	7294	PASS
173	174	0.00	2	1.0	946	PASS
174	95	50	100	86.2	94269	PASS
175	174	5	9	7.6	7162	PASS
176	174	95	101	96.9	91376	PASS
177	176	5	9	6.8	6258	PASS