

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081821\
 Data File : VY005737.D
 Acq On : 18 Aug 2021 14:12
 Operator : SY/MD
 Sample : M3430-04
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-4

Quant Time: Aug 19 05:17:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1764	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.697	114	2911	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	2239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	352	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	1071	59.980	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.960%
35) Dibromofluoromethane	7.734	113	500	30.679	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	61.360%
50) Toluene-d8	10.185	98	3532	51.673	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.340%
62) 4-Bromofluorobenzene	12.483	95	506	22.355	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	44.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081821\
Data File : VY005737.D
Acq On : 18 Aug 2021 14:12
Operator : SY/MD
Sample : M3430-04
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-4

Quant Time: Aug 19 05:17:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

