

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081821\
 Data File : VY005742.D
 Acq On : 18 Aug 2021 16:08
 Operator : SY/MD
 Sample : M3428-07
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-03-24.5-25.0

Quant Time: Aug 19 05:18:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.807	168	874	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	2506	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	3170	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	1809	204.475	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	408.940%#
35) Dibromofluoromethane	7.728	113	558	39.771	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.540%
50) Toluene-d8	10.179	98	3139	53.345	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.700%
62) 4-Bromofluorobenzene	12.483	95	1253	64.303	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	128.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081821\
Data File : VY005742.D
Acq On : 18 Aug 2021 16:08
Operator : SY/MD
Sample : M3428-07
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-03-24.5-25.0

Quant Time: Aug 19 05:18:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

