

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081821\
 Data File : VY005744.D
 Acq On : 18 Aug 2021 16:54
 Operator : SY/MD
 Sample : M3428-13
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-05-6.5-7.0

Quant Time: Aug 19 05:18:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	181867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	316806	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	330087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	167052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	113599	61.707	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.420%	
35) Dibromofluoromethane	7.728	113	96065	54.160	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.320%	
50) Toluene-d8	10.185	98	411218	55.280	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.560%	
62) 4-Bromofluorobenzene	12.483	95	143782	58.368	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 116.740%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081821\
Data File : VY005744.D
Acq On : 18 Aug 2021 16:54
Operator : SY/MD
Sample : M3428-13
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-05-6.5-7.0

Quant Time: Aug 19 05:18:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

