

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081821\
 Data File : VY005752.D
 Acq On : 18 Aug 2021 20:00
 Operator : SY/MD
 Sample : M3414-08
 Misc : 5.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-5B

Quant Time: Aug 19 05:19:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	11946	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	19991	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	22670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	10111	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	9652	79.819	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	159.640%
35) Dibromofluoromethane	7.728	113	7006	62.596	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	125.200%
50) Toluene-d8	10.185	98	26014	55.419	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.840%
62) 4-Bromofluorobenzene	12.483	95	9233	59.398	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081821\
Data File : VY005752.D
Acq On : 18 Aug 2021 20:00
Operator : SY/MD
Sample : M3414-08
Misc : 5.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-5B

Quant Time: Aug 19 05:19:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

