

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081822\
 Data File : VY010087.D
 Acq On : 18 Aug 2022 16:47
 Operator : KP/MD
 Sample : N4188-04RE
 Misc : 12.45g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S3RE

Quant Time: Aug 18 19:09:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 17 05:12:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	48276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	87384	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	87671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	39158	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	31881	66.048	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.100%
35) Dibromofluoromethane	7.710	113	30039	53.008	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.020%
50) Toluene-d8	10.179	98	105647	50.390	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.780%
62) 4-Bromofluorobenzene	12.477	95	34150	46.374	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081822\
Data File : VY010087.D
Acq On : 18 Aug 2022 16:47
Operator : KP/MD
Sample : N4188-04RE
Misc : 12.45g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S3RE

Quant Time: Aug 18 19:09:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 17 05:12:02 2022
Response via : Initial Calibration

