



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081822\
 Data File : VY010082.D
 Acq On : 18 Aug 2022 11:41
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081522S.M
 Title : SW846 8260
 Last Update : Wed Aug 17 05:12:02 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	7980	PASS
75	95	30	60	49.6	24629	PASS
95	95	100	100	100.0	49653	PASS
96	95	5	9	6.9	3429	PASS
173	174	0.00	2	1.5	659	PASS
174	95	50	100	88.9	44144	PASS
175	174	5	9	7.4	3260	PASS
176	174	95	101	96.4	42547	PASS
177	176	5	9	6.9	2953	PASS