



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081823\
Data File : VY015183.D
Acq On : 18 Aug 2023 08:52
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Title : SW846 8260
Last Update : Mon Aug 14 12:43:26 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.8	20115	PASS
75	95	30	60	53.4	54421	PASS
95	95	100	100	100.0	101837	PASS
96	95	5	9	6.8	6944	PASS
173	174	0.00	2	1.0	879	PASS
174	95	50	100	86.2	87784	PASS
175	174	5	9	7.7	6784	PASS
176	174	95	101	97.8	85877	PASS
177	176	5	9	6.8	5870	PASS