

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081921\
 Data File : VY005770.D
 Acq On : 19 Aug 2021 15:35
 Operator : SY/MD
 Sample : M3414-09
 Misc : 5.01G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-6A

Quant Time: Aug 20 06:56:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	49469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	77472	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	78293	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	37810	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	29036	57.985	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.980%
35) Dibromofluoromethane	7.728	113	24513	56.515	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.020%
50) Toluene-d8	10.185	98	100535	55.266	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.540%
62) 4-Bromofluorobenzene	12.483	95	32020	53.154	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.300%
Target Compounds						
16) Acetone	3.960	43	1241	9.830	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081921\
Data File : VY005770.D
Acq On : 19 Aug 2021 15:35
Operator : SY/MD
Sample : M3414-09
Misc : 5.01G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-6A

Quant Time: Aug 20 06:56:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

