

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081921\
 Data File : VY005773.D
 Acq On : 19 Aug 2021 16:44
 Operator : SY/MD
 Sample : M3418-12
 Misc : 10.26G/5ML/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SWCS-11-0205

Quant Time: Aug 20 06:57:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	204896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	357532	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	374114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	194920	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	135058	65.118	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.240%
35) Dibromofluoromethane	7.728	113	108577	54.241	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.480%
50) Toluene-d8	10.185	98	465344	55.430	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.860%
62) 4-Bromofluorobenzene	12.483	95	164573	59.198	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081921\
Data File : VY005773.D
Acq On : 19 Aug 2021 16:44
Operator : SY/MD
Sample : M3418-12
Misc : 10.26G/5ML/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SWCS-11-0205

Quant Time: Aug 20 06:57:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

