

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081921\
 Data File : VY005775.D
 Acq On : 19 Aug 2021 17:30
 Operator : SY/MD
 Sample : M3418-14
 Misc : 5.08G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SWCS-11-1015

Quant Time: Aug 20 06:58:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	203908	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	312849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	330334	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	172361	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	120110	58.191	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.380%
35) Dibromofluoromethane	7.728	113	110858	63.291	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	126.580%
50) Toluene-d8	10.185	98	404846	55.111	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.220%
62) 4-Bromofluorobenzene	12.483	95	145207	59.692	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081921\
Data File : VY005775.D
Acq On : 19 Aug 2021 17:30
Operator : SY/MD
Sample : M3418-14
Misc : 5.08G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SWCS-11-1015

Quant Time: Aug 20 06:58:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

