

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081921\
 Data File : VY005780.D
 Acq On : 19 Aug 2021 19:27
 Operator : SY/MD
 Sample : M3414-09
 Misc : 5.11G/5ML/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-6A

Quant Time: Aug 20 06:59:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	176719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	304403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	319494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	168336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	113891	63.668	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.340%	
35) Dibromofluoromethane	7.728	113	92798	54.450	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.900%	
50) Toluene-d8	10.185	98	396140	55.423	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.840%	
62) 4-Bromofluorobenzene	12.483	95	141205	59.657	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 119.320%	
Target Compounds						
16) Acetone	3.954	43	6036	13.383	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081921\
Data File : VY005780.D
Acq On : 19 Aug 2021 19:27
Operator : SY/MD
Sample : M3414-09
Misc : 5.11G/5ML/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-6A

Quant Time: Aug 20 06:59:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

