

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081922\
 Data File : VY010101.D
 Acq On : 19 Aug 2022 14:05
 Operator : KP/MD
 Sample : BLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BLK

Quant Time: Aug 22 02:39:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 22 02:37:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	188706	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	339114	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	320054	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	142261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	102214	66.129	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	132.260%
35) Dibromofluoromethane	7.710	113	104776	54.832	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.660%
50) Toluene-d8	10.179	98	393633	54.976	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.960%
62) 4-Bromofluorobenzene	12.477	95	126992	49.640	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.280%
Target Compounds						
2) Dichlorodifluoromethane	1.900	85	2350	1.671	ug/l	98
20) Methylene Chloride	4.704	84	40553	12.341	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081922\
Data File : VY010101.D
Acq On : 19 Aug 2022 14:05
Operator : KP/MD
Sample : BLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BLK

Quant Time: Aug 22 02:39:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 22 02:37:07 2022
Response via : Initial Calibration

