

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081922\
 Data File : VY010109.D
 Acq On : 19 Aug 2022 17:28
 Operator : KP/MD
 Sample : N4200-04
 Misc : 8.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S9

Quant Time: Aug 22 02:43:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 22 02:37:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	135791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	258619	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	259554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	125626	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	96751	86.986	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	173.980%#
35) Dibromofluoromethane	7.710	113	91219	62.596	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	125.200%
50) Toluene-d8	10.179	98	313648	57.439	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.880%
62) 4-Bromofluorobenzene	12.477	95	100997	51.767	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081922\
Data File : VY010109.D
Acq On : 19 Aug 2022 17:28
Operator : KP/MD
Sample : N4200-04
Misc : 8.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S9

Quant Time: Aug 22 02:43:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 22 02:37:07 2022
Response via : Initial Calibration

