

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081922\
 Data File : VY010116.D
 Acq On : 19 Aug 2022 20:13
 Operator : KP/MD
 Sample : N4217-08
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP3-0815

Quant Time: Aug 22 02:45:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 22 02:37:07 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	177868	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	338111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	335993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	154636	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	102534	70.378	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.760%	
35) Dibromofluoromethane	7.710	113	96401	50.599	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.200%	
50) Toluene-d8	10.173	98	353791	49.558	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.120%	
62) 4-Bromofluorobenzene	12.477	95	114578	44.920	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.840%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.942	59	5040	42.429	ug/l #	82
16) Acetone	3.942	43	23183	63.578	ug/l	93
17) Carbon Disulfide	4.174	76	5810	1.117	ug/l	95
20) Methylene Chloride	4.704	84	35072	10.533	ug/l	93
25) 2-Butanone	6.978	43	13029	22.213	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081922\
Data File : VY010116.D
Acq On : 19 Aug 2022 20:13
Operator : KP/MD
Sample : N4217-08
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP3-0815

Quant Time: Aug 22 02:45:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 22 02:37:07 2022
Response via : Initial Calibration

