


Tert butyl alcohol

Response Ratio



$$\text{Response} = 4.431\text{e-}002 * \text{Amt} - 9.265\text{e-}003$$


Coef of Det (r^2) = 0.992808 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y081922S.M

Calibration Table Last Updated: Mon Aug 22 02:37:07 2022

Naphthalene

Response Ratio



$$\text{Response} = 2.037\text{e}+000 * \text{Amt} - 1.583\text{e}-001$$


Coef of Det (r^2) = 0.998698 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y081922S.M

Calibration Table Last Updated: Mon Aug 22 02:37:07 2022

Methylene Chloride

Response Ratio




$$\text{Response} = 4.902\text{e-}001 * \text{Amt} + 9.391\text{e-}002$$

Coef of Det (r^2) = 0.998807 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y081922S.M

Calibration Table Last Updated: Mon Aug 22 02:37:07 2022

Acrolein



Response = $1.888 \times 10^{-2} \times \text{Amt} + 8.963 \times 10^{-3}$
 Coef of Det (r^2) = 0.997240 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA Y\methods\82Y081922S.M
 Calibration Table Last Updated: Mon Aug 22 02:37:07 2022