



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081922\
 Data File : VY010092.D
 Acq On : 19 Aug 2022 09:15
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Title : SW846 8260
 Last Update : Mon Aug 22 02:37:07 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	16800	PASS
75	95	30	60	47.7	49640	PASS
95	95	100	100	100.0	104075	PASS
96	95	5	9	6.6	6833	PASS
173	174	0.00	2	1.2	1141	PASS
174	95	50	100	90.9	94581	PASS
175	174	5	9	7.4	7035	PASS
176	174	95	101	95.7	90531	PASS
177	176	5	9	6.6	5977	PASS