



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081922\
 Data File : VY010118.D
 Acq On : 19 Aug 2022 21:22
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Title : SW846 8260
 Last Update : Mon Aug 22 02:37:07 2022



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	16912	PASS
75	95	30	60	50.0	46576	PASS
95	95	100	100	100.0	93243	PASS
96	95	5	9	7.0	6489	PASS
173	174	0.00	2	1.5	1217	PASS
174	95	50	100	87.8	81835	PASS
175	174	5	9	7.4	6024	PASS
176	174	95	101	96.0	78597	PASS
177	176	5	9	6.8	5310	PASS