

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081924\
 Data File : VY019114.D
 Acq On : 19 Aug 2024 14:01
 Operator : SY/MD
 Sample : P3615-07
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-04-081424

Quant Time: Aug 19 15:24:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	22674	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.604	114	33409	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	26041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	8652	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	8648	38.048	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	76.100%
35) Dibromofluoromethane	7.628	113	10264	46.986	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.980%
50) Toluene-d8	10.103	98	35735	45.458	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.920%
62) 4-Bromofluorobenzene	12.402	95	9217	34.394	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	68.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081924\
Data File : VY019114.D
Acq On : 19 Aug 2024 14:01
Operator : SY/MD
Sample : P3615-07
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-04-081424

Quant Time: Aug 19 15:24:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

