

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081924\
 Data File : VY019115.D
 Acq On : 19 Aug 2024 14:26
 Operator : SY/MD
 Sample : P3614-01RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-01-081424RE

Quant Time: Aug 19 15:24:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.695	168	7468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	11104	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	7142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	1755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	3790	50.627	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.260%
35) Dibromofluoromethane	7.622	113	3864	53.220	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.440%
50) Toluene-d8	10.103	98	11095	42.464	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.920%
62) 4-Bromofluorobenzene	12.414	95	2136	23.981	ug/l	0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	47.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081924\
Data File : VY019115.D
Acq On : 19 Aug 2024 14:26
Operator : SY/MD
Sample : P3614-01RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-01-081424RE

Quant Time: Aug 19 15:24:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

