

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
 Data File : VY005792.D
 Acq On : 20 Aug 2021 14:55
 Operator : SY/MD
 Sample : M3467-16
 Misc : 2.08G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SAMPLE-3-(171.00)-VOC-GRAB

Quant Time: Aug 23 01:07:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	178457	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	295319	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	276454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	103478	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	118589	65.648	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.300%
35) Dibromofluoromethane	7.728	113	93551	56.580	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.160%
50) Toluene-d8	10.185	98	378546	54.590	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.180%
62) 4-Bromofluorobenzene	12.483	95	105781	46.066	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
Data File : VY005792.D
Acq On : 20 Aug 2021 14:55
Operator : SY/MD
Sample : M3467-16
Misc : 2.08G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SAMPLE-3-(171.00)-VOC-GRAB

Quant Time: Aug 23 01:07:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

