

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
 Data File : VY005794.D
 Acq On : 20 Aug 2021 15:42
 Operator : SY/MD
 Sample : M3467-20
 Misc : 3.77G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SAMPLE-5-(184.50)-VOC-GRAB

Quant Time: Aug 23 01:08:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	181064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	298640	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	243603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	70885	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	113866	62.126	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.260%
35) Dibromofluoromethane	7.728	113	94226	56.355	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.700%
50) Toluene-d8	10.185	98	374756	53.443	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.880%
62) 4-Bromofluorobenzene	12.483	95	81326	35.022	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
Data File : VY005794.D
Acq On : 20 Aug 2021 15:42
Operator : SY/MD
Sample : M3467-20
Misc : 3.77G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SAMPLE-5-(184.50)-VOC-GRAB

Quant Time: Aug 23 01:08:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

