

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
 Data File : VY005795.D
 Acq On : 20 Aug 2021 16:05
 Operator : SY/MD
 Sample : M3467-22
 Misc : 4.67G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SAMPLE-6-(203.30)-VOC-GRAB

Quant Time: Aug 23 01:08:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	182066	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	305890	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	263614	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	81836	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	115411	62.623	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.240%
35) Dibromofluoromethane	7.728	113	95852	55.969	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.940%
50) Toluene-d8	10.185	98	385707	53.701	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.400%
62) 4-Bromofluorobenzene	12.483	95	90137	37.897	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
Data File : VY005795.D
Acq On : 20 Aug 2021 16:05
Operator : SY/MD
Sample : M3467-22
Misc : 4.67G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SAMPLE-6-(203.30)-VOC-GRAB

Quant Time: Aug 23 01:08:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

