

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
 Data File : VY005796.D
 Acq On : 20 Aug 2021 16:28
 Operator : SY/MD
 Sample : M3485-01
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-02-081921

Quant Time: Aug 23 01:08:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	186015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	319433	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	328794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	166381	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	108170	57.448	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.900%
35) Dibromofluoromethane	7.728	113	94750	52.979	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.960%
50) Toluene-d8	10.185	98	413364	55.111	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.220%
62) 4-Bromofluorobenzene	12.483	95	141534	56.983	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.960%
Target Compounds						
20) Methylene Chloride	4.722	84	72318	36.836	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
Data File : VY005796.D
Acq On : 20 Aug 2021 16:28
Operator : SY/MD
Sample : M3485-01
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-02-081921

Quant Time: Aug 23 01:08:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

