

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
 Data File : VY005797.D
 Acq On : 20 Aug 2021 16:51
 Operator : SY/MD
 Sample : M3465-01
 Misc : 5.14G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CON-A

Quant Time: Aug 23 01:08:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	179847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	309404	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	323278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	169287	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	111997	61.520	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.040%
35) Dibromofluoromethane	7.728	113	71313	41.167	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.340%
50) Toluene-d8	10.184	98	399040	54.926	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.860%
62) 4-Bromofluorobenzene	12.483	95	144281	59.972	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.940%
Target Compounds						
16) Acetone	3.960	43	19458	42.393	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
Data File : VY005797.D
Acq On : 20 Aug 2021 16:51
Operator : SY/MD
Sample : M3465-01
Misc : 5.14G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CON-A

Quant Time: Aug 23 01:08:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

