

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
 Data File : VY005800.D
 Acq On : 20 Aug 2021 18:01
 Operator : SY/MD
 Sample : M3451-01
 Misc : 4.68G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-0-5-COMP

Quant Time: Aug 23 01:09:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	190456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	325235	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	340418	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	177131	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	122341	63.459	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.920%	
35) Dibromofluoromethane	7.728	113	99486	54.635	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.280%	
50) Toluene-d8	10.185	98	428411	56.098	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 112.200%	
62) 4-Bromofluorobenzene	12.483	95	148898	58.878	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 117.760%	
Target Compounds						Qvalue
16) Acetone	3.960	43	5283	10.869	ug/l	91
20) Methylene Chloride	4.722	84	24526	7.585	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
Data File : VY005800.D
Acq On : 20 Aug 2021 18:01
Operator : SY/MD
Sample : M3451-01
Misc : 4.68G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-0-5-COMP

Quant Time: Aug 23 01:09:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

