

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
 Data File : VY005804.D
 Acq On : 20 Aug 2021 19:34
 Operator : SY/MD
 Sample : M3463-23
 Misc : 5.79G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SWCS-19-0204

Quant Time: Aug 23 01:10:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	183485	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	314473	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	292919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	107999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	113532	61.127	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.260%	
35) Dibromofluoromethane	7.722	113	94495	53.670	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.340%	
50) Toluene-d8	10.185	98	395594	53.574	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.140%	
62) 4-Bromofluorobenzene	12.483	95	113038	46.228	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.460%	
Target Compounds						Qvalue
16) Acetone	3.972	43	10569	22.570	ug/l	100
44) Trichloroethene	8.941	130	19459	8.441	ug/l	94
64) Tetrachloroethene	10.721	164	4252	1.907	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
Data File : VY005804.D
Acq On : 20 Aug 2021 19:34
Operator : SY/MD
Sample : M3463-23
Misc : 5.79G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SWCS-19-0204

Quant Time: Aug 23 01:10:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

