

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
 Data File : VY005805.D
 Acq On : 20 Aug 2021 19:57
 Operator : SY/MD
 Sample : M3463-25
 Misc : 7.96G/5ML/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SWCS-23-0002

Quant Time: Aug 23 01:11:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	191917	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	332395	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	328692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	140787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	123162	63.398	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 126.800%			
35) Dibromofluoromethane	7.728	113	99852	53.655	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 107.320%			
50) Toluene-d8	10.185	98	421883	54.054	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 108.100%			
62) 4-Bromofluorobenzene	12.483	95	135557	52.448	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 104.900%			
Target Compounds						Qvalue
16) Acetone	3.966	43	27062	55.252	ug/l	92
65) Chlorobenzene	11.520	112	22836	3.452	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082021\
Data File : VY005805.D
Acq On : 20 Aug 2021 19:57
Operator : SY/MD
Sample : M3463-25
Misc : 7.96G/5ML/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SWCS-23-0002

Quant Time: Aug 23 01:11:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

