

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082123\
 Data File : VY015226.D
 Acq On : 21 Aug 2023 18:32
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Aug 22 01:57:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	139063	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	249167	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	247632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	123360	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	91306	54.696	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery	= 109.400%		
35) Dibromofluoromethane	7.710	113	77583	48.399	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery	= 96.800%		
50) Toluene-d8	10.173	98	296585	49.703	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery	= 99.400%		
62) 4-Bromofluorobenzene	12.477	95	113087	51.887	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery	= 103.780%		
Target Compounds						Qvalue
20) Methylene Chloride	4.698	84	43088	31.341	ug/l	93
52) Toluene	10.240	92	5148	1.457	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082123\
Data File : VY015226.D
Acq On : 21 Aug 2023 18:32
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Aug 22 01:57:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

