

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082123\
 Data File : VY015227.D
 Acq On : 21 Aug 2023 18:56
 Operator : SY/MD
 Sample : 04091-01
 Misc : 4.98g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-A

Quant Time: Aug 22 01:57:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	126320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	223813	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	228074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	116473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	86021	56.729	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.460%	
35) Dibromofluoromethane	7.710	113	72607	50.426	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.860%	
50) Toluene-d8	10.173	98	273407	51.009	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.020%	
62) 4-Bromofluorobenzene	12.477	95	106662	54.483	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 108.960%	
Target Compounds						Qvalue
20) Methylene Chloride	4.704	84	25148	18.447	ug/l	91
52) Toluene	10.240	92	4648	1.464	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082123\
Data File : VY015227.D
Acq On : 21 Aug 2023 18:56
Operator : SY/MD
Sample : 04091-01
Misc : 4.98g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-A

Quant Time: Aug 22 01:57:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

