

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082123\
 Data File : VY015221.D
 Acq On : 21 Aug 2023 16:28
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Aug 22 01:55:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.783	168	138513	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	245539	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	248254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124048	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	93353	56.145	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.280%
35) Dibromofluoromethane	7.710	113	78727	49.839	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.680%
50) Toluene-d8	10.179	98	295792	50.302	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.600%
62) 4-Bromofluorobenzene	12.477	95	115755	53.895	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.800%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.704	84	47119	34.873	ug/l	88
52) Toluene	10.240	92	14657	4.209	ug/l	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082123\
Data File : VY015221.D
Acq On : 21 Aug 2023 16:28
Operator : SY/MD
Sample : VIBLK
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Aug 22 01:55:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY015221.D
Acq: 21 Aug 2023 16:28




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion:168 Resp: 138513
Ion Ratio Lower Upper
168 100
99 55.4 43.4 65.0



#20
Methylene Chloride
Concen: 34.873 ug/l
RT: 4.704 min Scan# 473
Delta R.T. -0.006 min
Lab File: VY015221.D
Acq: 21 Aug 2023 16:28

Tgt Ion: 84 Resp: 47119
Ion Ratio Lower Upper
84 100
49 116.2 106.9 160.3
51 35.5 33.2 49.8
86 69.0 51.0 76.6





#33

1,2-Dichloroethane-d4

Concen: 56.145 ug/l

RT: 8.137 min Scan# 1036

Delta R.T. 0.000 min

Lab File: VY015221.D

Acq: 21 Aug 2023 16:28

Instrument :

MSVOA_Y

ClientSampleId :





VIBLK

Tgt Ion: 65 Resp: 93353

Ion Ratio Lower Upper

65 100

67 51.4 0.0 101.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY015221.D

Acq: 21 Aug 2023 16:28




Tgt Ion: 114 Resp: 245539

Ion Ratio Lower Upper

114 100

63 19.4 0.0 42.4

88 15.0 0.0 30.6









#35
Dibromofluoromethane
Concen: 49.839 ug/l
RT: 7.710 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY015221.D
Acq: 21 Aug 2023 16:28




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion:113 Resp: 78727
Ion Ratio Lower Upper
113 100
111 102.6 81.3 121.9
192 19.4 16.9 25.3



#50
Toluene-d8
Concen: 50.302 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. 0.000 min
Lab File: VY015221.D
Acq: 21 Aug 2023 16:28

Tgt Ion: 98 Resp: 295792
Ion Ratio Lower Upper
98 100
100 65.5 50.1 75.1









#52
Toluene
Concen: 4.209 ug/l
RT: 10.240 min Scan# 1381
Delta R.T. 0.000 min
Lab File: VY015221.D
Acq: 21 Aug 2023 16:28




Instrument :
MSVOA_Y
ClientSampleId :
VIBLK


Tgt Ion: 92 Resp: 14657
Ion Ratio Lower Upper
92 100
91 168.6 143.8 215.6



#62
4-Bromofluorobenzene
Concen: 53.895 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.000 min
Lab File: VY015221.D
Acq: 21 Aug 2023 16:28

Tgt Ion: 95 Resp: 115755
Ion Ratio Lower Upper
95 100
174 84.5 0.0 178.2
176 80.7 0.0 173.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 117
Delta R.T. 0.000 min
Lab File: VY015221.D
Acq: 21 Aug 2023 16:28

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Tgt Ion:117 Resp: 248254
Ion Ratio Lower Upper
117 100
82 54.9 43.4 65.0
119 31.1 26.3 39.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. 0.000 min
Lab File: VY015221.D
Acq: 21 Aug 2023 16:28

Tgt Ion:152 Resp: 124048
Ion Ratio Lower Upper
152 100
115 57.6 28.8 86.5
150 159.6 0.0 348.8

