

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082223\
 Data File : VY015242.D
 Acq On : 22 Aug 2023 14:17
 Operator : SY/MD
 Sample : 04086-16
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-PP18B

Quant Time: Aug 22 15:13:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	84819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	138099	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	122927	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	56670	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	55118	54.134	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.260%
35) Dibromofluoromethane	7.715	113	43642	49.122	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.240%
50) Toluene-d8	10.178	98	163161	49.334	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.483	95	52229	43.237	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.480%
Target Compounds						
20) Methylene Chloride	4.710	84	9250	7.967	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082223\
Data File : VY015242.D
Acq On : 22 Aug 2023 14:17
Operator : SY/MD
Sample : 04086-16
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-PP18B

Quant Time: Aug 22 15:13:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

