

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082224\
 Data File : VY019233.D
 Acq On : 22 Aug 2024 20:23
 Operator : SY/MD
 Sample : P3703-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-8-21-2024

Quant Time: Aug 23 01:57:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	76918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	136265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	125785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	46770	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	39010	50.594	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.180%
35) Dibromofluoromethane	7.628	113	43557	48.887	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.780%
50) Toluene-d8	10.103	98	154008	48.032	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.060%
62) 4-Bromofluorobenzene	12.408	95	46113	42.188	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	84.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082224\
Data File : VY019233.D
Acq On : 22 Aug 2024 20:23
Operator : SY/MD
Sample : P3703-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-8-21-2024

Quant Time: Aug 23 01:57:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

