

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082322\
 Data File : VY010166.D
 Acq On : 23 Aug 2022 15:35
 Operator : KP/MD
 Sample : N4200-04RE
 Misc : 7.32g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S9RE

Quant Time: Aug 24 04:33:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 22 02:37:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	81951	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	156020	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	347964	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	141024	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	48152	71.734	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 143.460%	
35) Dibromofluoromethane	7.710	113	44886	51.056	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.120%	
50) Toluene-d8	10.173	98	162530	49.338	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.680%	
62) 4-Bromofluorobenzene	12.477	95	103739	88.138	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 176.280%#	
Target Compounds						Qvalue
11) Tert butyl alcohol	4.954	59	2185	40.541	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082322\
Data File : VY010166.D
Acq On : 23 Aug 2022 15:35
Operator : KP/MD
Sample : N4200-04RE
Misc : 7.32g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S9RE

Quant Time: Aug 24 04:33:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 22 02:37:07 2022
Response via : Initial Calibration

