

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082322\
 Data File : VY010175.D
 Acq On : 23 Aug 2022 19:08
 Operator : KP/MD
 Sample : N4275-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-081922

Quant Time: Aug 24 04:34:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 22 02:37:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	683	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	1258	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1156	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	473	84.549	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 169.100%#	
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	= 0.000%#	
50) Toluene-d8	10.179	98	1256	47.286	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 94.580%	
62) 4-Bromofluorobenzene	0.000	95	0	0.000	ug/l	
Spiked Amount	50.000	Range	25 - 144	Recovery	= 0.000%#	
Target Compounds						Qvalue
3) Chloromethane	2.119	50	33	5.762	ug/l	# 43

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082322\
Data File : VY010175.D
Acq On : 23 Aug 2022 19:08
Operator : KP/MD
Sample : N4275-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-081922

Quant Time: Aug 24 04:34:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 22 02:37:07 2022
Response via : Initial Calibration

