

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082322\
 Data File : VY010177.D
 Acq On : 23 Aug 2022 19:55
 Operator : KP/MD
 Sample : N4292-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-082222

Quant Time: Aug 24 04:35:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 22 02:37:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	163073	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	305978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	307385	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	141599	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	94202	70.525	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.060%
35) Dibromofluoromethane	7.710	113	84870	49.225	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.440%
50) Toluene-d8	10.179	98	313089	48.462	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.920%
62) 4-Bromofluorobenzene	12.483	95	100213	43.415	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082322\
Data File : VY010177.D
Acq On : 23 Aug 2022 19:55
Operator : KP/MD
Sample : N4292-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-082222

Quant Time: Aug 24 04:35:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 22 02:37:07 2022
Response via : Initial Calibration

