

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082322\
 Data File : VY010183.D
 Acq On : 23 Aug 2022 22:16
 Operator : KP/MD
 Sample : N4188-25
 Misc : 10.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S6

Quant Time: Aug 24 04:36:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 22 02:37:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	170610	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	319304	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	324197	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	150796	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	99915	71.498	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.000%
35) Dibromofluoromethane	7.710	113	91504	50.858	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.720%
50) Toluene-d8	10.179	98	332607	49.335	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.660%
62) 4-Bromofluorobenzene	12.477	95	106997	44.419	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	88.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082322\
Data File : VY010183.D
Acq On : 23 Aug 2022 22:16
Operator : KP/MD
Sample : N4188-25
Misc : 10.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S6

Quant Time: Aug 24 04:36:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081922S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 22 02:37:07 2022
Response via : Initial Calibration

