

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
 Data File : VY015292.D
 Acq On : 23 Aug 2023 19:43
 Operator : SY/MD
 Sample : 04119-03
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB21153

Quant Time: Aug 24 02:36:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1600	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	2754	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	3147	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1445	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	1375	71.590	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 143.180%	
35) Dibromofluoromethane	7.710	113	377	21.278	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 42.560%#	
50) Toluene-d8	10.179	98	3618	54.856	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 109.720%	
62) 4-Bromofluorobenzene	12.483	95	1861	77.253	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 154.500%#	
Target Compounds						Qvalue
3) Chloromethane	2.113	50	270	26.459	ug/l #	79
20) Methylene Chloride	4.704	84	609	39.581	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
Data File : VY015292.D
Acq On : 23 Aug 2023 19:43
Operator : SY/MD
Sample : 04119-03
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB21153

Quant Time: Aug 24 02:36:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

