

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
 Data File : VY015294.D
 Acq On : 23 Aug 2023 20:30
 Operator : SY/MD
 Sample : 04116-03
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-2

Quant Time: Aug 24 02:37:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	124148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	207612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	187674	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	91189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	81705	54.825	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.660%	
35) Dibromofluoromethane	7.710	113	66039	49.444	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.880%	
50) Toluene-d8	10.179	98	244525	49.180	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.360%	
62) 4-Bromofluorobenzene	12.483	95	81851	45.072	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 90.140%	
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	7143	14.672	ug/l	96
17) Carbon Disulfide	4.192	76	5062	3.330	ug/l #	92
20) Methylene Chloride	4.710	84	30398	23.775	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
Data File : VY015294.D
Acq On : 23 Aug 2023 20:30
Operator : SY/MD
Sample : 04116-03
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-2

Quant Time: Aug 24 02:37:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

