

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
 Data File : VY015295.D
 Acq On : 23 Aug 2023 20:54
 Operator : SY/MD
 Sample : 04116-05
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3

Quant Time: Aug 24 02:37:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	122779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	205989	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	185805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	88952	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	82152	55.740	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.480%
35) Dibromofluoromethane	7.716	113	65973	49.784	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.560%
50) Toluene-d8	10.179	98	240839	48.821	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.640%
62) 4-Bromofluorobenzene	12.483	95	81090	45.005	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.000%
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	11348	23.569	ug/l	96
17) Carbon Disulfide	4.186	76	6412	4.265	ug/l #	95
20) Methylene Chloride	4.710	84	20633	14.834	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
Data File : VY015295.D
Acq On : 23 Aug 2023 20:54
Operator : SY/MD
Sample : 04116-05
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3

Quant Time: Aug 24 02:37:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

