

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
 Data File : VY015296.D
 Acq On : 23 Aug 2023 21:17
 Operator : SY/MD
 Sample : 04116-07
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-4

Quant Time: Aug 24 02:37:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	125093	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	208248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	188969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	90924	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	83551	55.640	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.280%
35) Dibromofluoromethane	7.716	113	66752	49.825	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.640%
50) Toluene-d8	10.179	98	247757	49.678	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.360%
62) 4-Bromofluorobenzene	12.477	95	82566	45.327	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.660%
Target Compounds						
						Qvalue
17) Carbon Disulfide	4.180	76	2465	1.609	ug/l	99
20) Methylene Chloride	4.710	84	34101	27.006	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
Data File : VY015296.D
Acq On : 23 Aug 2023 21:17
Operator : SY/MD
Sample : 04116-07
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-4

Quant Time: Aug 24 02:37:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

