

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
 Data File : VY015303.D
 Acq On : 24 Aug 2023 00:01
 Operator : SY/MD
 Sample : 04110-24
 Misc : 4.79g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-20(2-3)

Quant Time: Aug 24 02:40:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	122822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	201467	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	170874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	66333	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	78150	53.006	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.020%	
35) Dibromofluoromethane	7.716	113	65038	50.180	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.360%	
50) Toluene-d8	10.179	98	236574	49.032	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.060%	
62) 4-Bromofluorobenzene	12.477	95	66703	37.851	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 75.700%	
Target Compounds						
17) Carbon Disulfide	4.192	76	5565	3.700	ug/l	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
Data File : VY015303.D
Acq On : 24 Aug 2023 00:01
Operator : SY/MD
Sample : 04110-24
Misc : 4.79g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-20(2-3)

Quant Time: Aug 24 02:40:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

