

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
 Data File : VY015305.D
 Acq On : 24 Aug 2023 00:48
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 24 02:41:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	63	0.00
2 T	Dichlorodifluoromethane	0.253	0.196	22.5	57	0.00
3 P	Chloromethane	0.347	0.290	16.4	60	0.00
4 C	Vinyl Chloride	0.398	0.362	9.0#	60	0.00
5 T	Bromomethane	0.291	0.329	-13.1	71	0.00
6 T	Chloroethane	0.289	0.293	-1.4	64	0.00
7 T	Trichlorofluoromethane	0.670	0.660	1.5	65	0.00
8 T	Diethyl Ether	0.263	0.286	-8.7	71	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.417	0.414	0.7	65	0.01
10 T	Methyl Iodide	0.409	0.356	13.0	52	0.00
11 T	Tert butyl alcohol	0.116	0.090	22.4	58	0.01
12 CM	1,1-Dichloroethene	0.348	0.336	3.4#	62	0.00
13 T	Acrolein	0.056	0.019	66.1#	34#	0.00
14 T	Allyl chloride	0.617	0.611	1.0	63	0.00
15 T	Acrylonitrile	0.178	0.208	-16.9	74	0.00
16 T	Acetone	0.196	0.171	12.8	53	0.00
17 T	Carbon Disulfide	0.612	0.523	14.5	53	0.00
18 T	Methyl Acetate	0.719	0.829	-15.3	73	0.01
19 T	Methyl tert-butyl Ether	1.479	1.609	-8.8	69	0.01
20 T	Methylene Chloride	0.570	0.587	-3.0	77	0.00
21 T	trans-1,2-Dichloroethene	0.392	0.397	-1.3	65	0.00
22 T	Diisopropyl ether	1.539	1.726	-12.2	71	0.00
23 T	Vinyl Acetate	0.918	0.983	-7.1	67	0.00
24 P	1,1-Dichloroethane	0.833	0.900	-8.0	69	0.00
25 T	2-Butanone	0.278	0.301	-8.3	68	0.00
26 T	2,2-Dichloropropane	0.803	0.706	12.1	58	0.00
27 T	cis-1,2-Dichloroethene	0.536	0.580	-8.2	70	0.00
28 T	Bromochloromethane	0.383	0.391	-2.1	66	0.00
29 T	Tetrahydrofuran	0.166	0.190	-14.5	73	0.00
30 C	Chloroform	0.934	1.028	-10.1#	71	0.00
31 T	Cyclohexane	0.597	0.542	9.2	62	0.01
32 T	1,1,1-Trichloroethane	0.795	0.829	-4.3	67	0.00
33 S	1,2-Dichloroethane-d4	0.600	0.639	-6.5	66	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	64	0.00
35 S	Dibromofluoromethane	0.322	0.337	-4.7	66	0.00
36 T	1,1-Dichloropropene	0.356	0.353	0.8	65	0.00
37 T	Ethyl Acetate	0.332	0.369	-11.1	74	0.00
38 T	Carbon Tetrachloride	0.421	0.414	1.7	65	0.00
39 T	Methylcyclohexane	0.399	0.372	6.8	61	0.00
40 TM	Benzene	1.076	1.136	-5.6	69	0.00
41 T	Methacrylonitrile	0.171	0.165	3.5	59	0.00
42 TM	1,2-Dichloroethane	0.391	0.423	-8.2	70	0.00
43 T	Isopropyl Acetate	0.604	0.656	-8.6	71	0.00
44 TM	Trichloroethene	0.310	0.320	-3.2	68	0.00
45 C	1,2-Dichloropropane	0.303	0.331	-9.2#	72	0.00
46 T	Dibromomethane	0.191	0.209	-9.4	70	0.00
47 T	Bromodichloromethane	0.457	0.505	-10.5	72	0.00
48 T	Methyl methacrylate	0.258	0.298	-15.5	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
 Data File : VY015305.D
 Acq On : 24 Aug 2023 00:48
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 24 02:41:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 14 12:43:26 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.004	0.0	66	0.00
50 S	Toluene-d8	1.197	1.215	-1.5	64	0.00
51 T	4-Methyl-2-Pentanone	0.350	0.408	-16.6	75	0.00
52 CM	Toluene	0.709	0.731	-3.1#	67	0.00
53 T	t-1,3-Dichloropropene	0.480	0.477	0.6	65	0.00
54 T	cis-1,3-Dichloropropene	0.505	0.522	-3.4	67	0.00
55 T	1,1,2-Trichloroethane	0.291	0.320	-10.0	72	0.00
56 T	Ethyl methacrylate	0.426	0.465	-9.2	69	0.00
57 T	1,3-Dichloropropane	0.475	0.522	-9.9	72	0.00
58 T	2-Chloroethyl Vinyl ether	0.198	0.170	14.1	57	0.00
59 T	2-Hexanone	0.259	0.290	-12.0	72	0.00
60 T	Dibromochloromethane	0.347	0.376	-8.4	70	0.00
61 T	1,2-Dibromoethane	0.276	0.289	-4.7	68	0.00
62 S	4-Bromofluorobenzene	0.437	0.435	0.5	64	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	65	0.00
64 T	Tetrachloroethene	0.348	0.392	-12.6	75	0.00
65 PM	Chlorobenzene	0.911	0.933	-2.4	68	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.391	-5.7	70	0.00
67 C	Ethyl Benzene	1.596	1.614	-1.1#	68	0.00
68 T	m/p-Xylenes	0.606	0.610	-0.7	67	0.00
69 T	o-Xylene	0.604	0.611	-1.2	67	0.00
70 T	Styrene	1.045	1.099	-5.2	69	0.00
71 P	Bromoform	0.268	0.280	-4.5	69	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
73 T	Isopropylbenzene	3.215	3.235	-0.6	68	0.00
74 T	N-amyl acetate	1.161	1.222	-5.3	68	0.00
75 P	1,1,2,2-Tetrachloroethane	0.837	0.919	-9.8	72	0.00
76 T	1,2,3-Trichloropropane	0.607	0.708	-16.6	79	0.00
77 T	Bromobenzene	0.784	0.812	-3.6	70	0.00
78 T	n-propylbenzene	3.818	3.892	-1.9	69	0.00
79 T	2-Chlorotoluene	2.203	2.261	-2.6	69	0.00
80 T	1,3,5-Trimethylbenzene	2.680	2.725	-1.7	68	0.00
81 T	trans-1,4-Dichloro-2-butene	0.277	0.259	6.5	60	0.00
82 T	4-Chlorotoluene	2.275	2.333	-2.5	68	0.00
83 T	tert-Butylbenzene	2.451	2.504	-2.2	68	0.00
84 T	1,2,4-Trimethylbenzene	2.678	2.743	-2.4	68	0.00
85 T	sec-Butylbenzene	3.592	3.685	-2.6	69	0.00
86 T	p-Isopropyltoluene	2.978	3.016	-1.3	67	0.00
87 T	1,3-Dichlorobenzene	1.554	1.615	-3.9	69	0.00
88 T	1,4-Dichlorobenzene	1.559	1.620	-3.9	69	0.00
89 T	n-Butylbenzene	2.808	2.808	0.0	66	0.00
90 T	Hexachloroethane	0.591	0.589	0.3	67	0.00
91 T	1,2-Dichlorobenzene	1.452	1.553	-7.0	71	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.162	0.175	-8.0	71	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.941	-1.0	67	0.00
94 T	Hexachlorobutadiene	0.507	0.489	3.6	66	0.00
95 T	Naphthalene	2.322	2.434	-4.8	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082323\
Data File : VY015305.D
Acq On : 24 Aug 2023 00:48
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Aug 24 02:41:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081423S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 14 12:43:26 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.842	0.876	-4.0	69	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6