

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082324\
 Data File : VY019255.D
 Acq On : 23 Aug 2024 14:16
 Operator : SY/MD
 Sample : P3663-03
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 914-J-SD-0-0.5-081624

Quant Time: Aug 23 22:57:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	79122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	145461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	125592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	44065	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	40762	51.393	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.780%
35) Dibromofluoromethane	7.628	113	45764	48.116	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.240%
50) Toluene-d8	10.103	98	161642	47.226	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.460%
62) 4-Bromofluorobenzene	12.408	95	45360	38.876	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	77.760%
Target Compounds						
17) Carbon Disulfide	4.086	76	3831	1.721	ug/l	Qvalue # 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082324\
Data File : VY019255.D
Acq On : 23 Aug 2024 14:16
Operator : SY/MD
Sample : P3663-03
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
914-J-SD-0-0.5-081624

Quant Time: Aug 23 22:57:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

