

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082324\
 Data File : VY019256.D
 Acq On : 23 Aug 2024 14:40
 Operator : SY/MD
 Sample : P3663-04
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 914-J-SD-0.5-1-081624

Quant Time: Aug 23 22:57:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	80419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	146211	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	131005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	43355	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	41751	51.791	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.580%
35) Dibromofluoromethane	7.622	113	46323	48.454	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.900%
50) Toluene-d8	10.103	98	165009	47.963	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.920%
62) 4-Bromofluorobenzene	12.408	95	46324	39.498	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	79.000%
Target Compounds						
16) Acetone	3.854	43	2761	18.468	ug/l	Qvalue # 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082324\
Data File : VY019256.D
Acq On : 23 Aug 2024 14:40
Operator : SY/MD
Sample : P3663-04
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
914-J-SD-0.5-1-081624

Quant Time: Aug 23 22:57:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

