

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082421\
 Data File : VY005817.D
 Acq On : 24 Aug 2021 13:14
 Operator : SY/MD
 Sample : M3463-07
 Misc : 3.21G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SWCS-20-0306

Quant Time: Aug 25 04:29:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	179529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	291154	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	250677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	82735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	101562	55.887	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.780%
35) Dibromofluoromethane	7.722	113	86469	53.045	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.100%
50) Toluene-d8	10.185	98	358304	52.410	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.820%
62) 4-Bromofluorobenzene	12.483	95	87882	38.819	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.640%
Target Compounds						
17) Carbon Disulfide	4.204	76	9966	2.090	ug/l	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082421\
Data File : VY005817.D
Acq On : 24 Aug 2021 13:14
Operator : SY/MD
Sample : M3463-07
Misc : 3.21G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SWCS-20-0306

Quant Time: Aug 25 04:29:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

