

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082421\
 Data File : VY005818.D
 Acq On : 24 Aug 2021 13:37
 Operator : SY/MD
 Sample : M3463-13
 Misc : 6.19G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SWCS-16-0002

Quant Time: Aug 25 04:29:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	197327	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	336178	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	329064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	151624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	118148	59.150	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.300%	
35) Dibromofluoromethane	7.728	113	99258	52.736	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.480%	
50) Toluene-d8	10.185	98	425213	53.867	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.740%	
62) 4-Bromofluorobenzene	12.483	95	145431	55.635	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 111.280%	
Target Compounds						
						Qvalue
17) Carbon Disulfide	4.204	76	6801	1.297	ug/l	# 89
64) Tetrachloroethene	10.721	164	8239	3.289	ug/l	95
65) Chlorobenzene	11.520	112	9351	1.412	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082421\
Data File : VY005818.D
Acq On : 24 Aug 2021 13:37
Operator : SY/MD
Sample : M3463-13
Misc : 6.19G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SWCS-16-0002

Quant Time: Aug 25 04:29:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

