

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082421\
 Data File : VY005820.D
 Acq On : 24 Aug 2021 14:47
 Operator : SY/MD
 Sample : M3463-19RE
 Misc : 3.80G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SWCS-18-0204RE

Quant Time: Aug 25 04:29:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 03 04:52:05 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	229258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	391030	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	291532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	102633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	131834	56.809	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.620%
35) Dibromofluoromethane	7.722	113	115808	52.898	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.800%
50) Toluene-d8	10.185	98	417272	45.446	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.900%
62) 4-Bromofluorobenzene	12.483	95	112296	36.933	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	73.860%
Target Compounds						Qvalue
4) Vinyl Chloride	2.259	62	5871	2.179	ug/l	# 88
16) Acetone	3.960	43	47914	81.891	ug/l	93
17) Carbon Disulfide	4.204	76	18364	3.015	ug/l	100
86) p-Isopropyltoluene	13.373	119	16103	2.437	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082421\
Data File : VY005820.D
Acq On : 24 Aug 2021 14:47
Operator : SY/MD
Sample : M3463-19RE
Misc : 3.80G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SWCS-18-0204RE

Quant Time: Aug 25 04:29:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y080221S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 03 04:52:05 2021
Response via : Initial Calibration

